



Draw Code	MIN	MAX
I.D	.10.0	10.1
W.T	0.024	0.035
O.L	1000.0	1010.0
I.L		
W.L		
W.W		

Confidential Property of:

Lamina Dielectrics Ltd, Daux Rd,  
Billingshurst, RH14 9SJ

Tel: +44(0) 1403 783131

Fax: +44(0) 1402 782237

Email: technical@lamina.uk.com

<b>CUSTOMER:</b>	Herbrew University	
<b>PROJECT:</b>		
<b>Drawn by:</b>	PMH	5.01.14
<b>Approved by:</b>	PR	5.01.14
<b>Material:</b>	Mylar EI 0.012 mm thick coated with 1,000 Angstroms of Aluminium	

All dimensions in mm:

**Tolerances**

Linear: Angular:

X +/- 0.5mm +/- 0.5°

X.X +/- 0.1mm

X.XX +/- 0.05mm

Unless otherwise stated on drawing

**DEBURR AND REMOVE ALL SHARP  
EDGES UNLESS STATED OTHERWISE**

**Notes:**

Two layer tube wound with Aluminium on  
both the inside and outside surfaces.

**TITLE:**

Polyester Aluminium coated detector  
straw

**Drawing No.:** 51.9267.02

**Issue:** 1 **Date:** 5.01.14

DWG SIZE A4